

American National Standard

*for Methods of Nuclear Material Control –
Measurement Control Program –
Nuclear Materials
Analytical Chemistry Laboratory*



ANSI[®]
N15.51-2007
Revision of
ANSI N15.51-1990 (R2006)

American National Standard
for Methods of
Nuclear Material Control –

Measurement Control Program –
Nuclear Materials
Analytical Chemistry Laboratory

Secretariat

Institute of Nuclear Materials Management

Approved October 31, 2007

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Published by

**American National Standards Institute, Inc.
25 West 43rd Street, New York, NY 10036**

Copyright © 2008 by American National Standards Institute, Inc.
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Page
Foreword	iii
0 Introduction	1
1 Scope and purpose	1
2 Normative references	2
3 Definitions	2
4 Technical Aspects of a Measurement Control Program	5
5 Administrative Aspects of a Measurement Control Program	10
Annexes	
A Measurement methods procedure requirements	13
B Target values for uncertainty components in fissile isotope and element assay	14
C Target values for random uncertainties in sampling and element assay of nuclear materials	15
D Target values and measurement system performance	16
E Method development and qualification report	17
F Sampling plan	18
G Packaging instruction requirements	19
H Identification and labeling	20
J Considerations for data records	21
K Storage of samples	22
L Considerations for laboratory documentation	23
M Program records considerations	24
N Measurement request and report form considerations	25
P Sample, laboratory, and data records requirements	26
Q Example of control charts	27
R Management of a measurement control program	29
S Essentials of a measurement control program	30
T Functional responsibilities for a measurement control program	31
U Training criteria for laboratory personnel	32
V Recommended qualification program for analysts	33
W Review and audit of measurement control	34

	Page
X Documentation	37
Y Response to out-of-control situations	38
Z Objectives/purposes of Interlaboratory Comparison Programs.....	39
AA Bibliography.....	40

Foreword (This foreword is not part of American National Standard ANSI N15.51-2007.)

This standard was revised under the procedures of the American National Standards Institute by Accredited Standards Committee N15 on Methods of Nuclear Material Control. The secretariat of N15 is held by the Institute of Nuclear Materials Management (INMM). Committee N15 has the following scope:

Standards for the protection, control, and accounting of special nuclear materials in all phases of the nuclear fuel cycle, including analytical procedures where necessary and special to this purpose, except that physical protection of special nuclear materials within a nuclear power plant is not included.

The Institute of Nuclear Materials Management has long recognized the importance of measurement quality for maintaining adequate protection, control, and accountability (MPC&A) for nuclear materials. Because of this commitment to quality, Writing Group INMM-5 on Measurement Control was established. The Writing Group issued a generic standard covering the general principles of good and acceptable measurement control, N15.41 "Derivation of Measurement Control Programs - General Principles." Additional standards were prepared to provide guidance for determining and monitoring the quality of various types of measurements made by the MPC&A community. This standard provides the principal elements of a measurement control program for an analytical chemistry laboratory supporting nuclear fuel cycle activities.

There are 25 annexes in this standard. Annexes A, F, G, H, K, P, R, S, T, U, W, and Y are normative and form part of the requirements of this standard. Annexes B, C, D, E, J, L, M, N, Q, V, X, Z, and AA are informative and are included for information only. (In accordance with the ISO Directives Part 3, there are no annexes designated with the letters 'I' or 'O'.)

Suggestions for improvement of the standard will be welcome. They should be sent to the Institute of Nuclear Materials Management, 60 Revere Drive, Suite 500, Northbrook, IL 60062. (Additional information about the INMM may be found at <http://www.inmm.org>.)

This standard was prepared by Committee N15 following ANSI requirements for due process and for obtaining consensus. N15 Committee approval of the standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the N15 Committee had the following members:

- Carrie Mathews, Chair
(Pacific Northwest National Laboratory)
- Melanie May, Vice-Chair
(U.S. Department of Energy)
- Lynne Preston, Secretary
(U.S. Department of Energy)

<i>Organization Represented</i>	<i>Name of Representative</i>
American Nuclear Society (ANS)	Patricia Schroeder
American Society of Industrial Security (ASIS).....	Robert D. Hulshouser
American Society for Quality Control (ASQC)	Chuck Moseley
American Society for Testing and Materials (ASTM).....	Charles Pietri
AREVA-NP	Marilyn Law
BWXT Lynchburg	Matthew Suwala
Global Nuclear Fuels	Rose Martyn
Institute of Nuclear Materials Management (INMM)	Nancy Jo Nicholas
Los Alamos National Laboratory (LANL)	James Sprinkle
National Institute of Standards and Technology (NIST)	James M. Adams

<i>Organization Represented</i>	<i>Name of Representative</i>
New Brunswick Laboratory (NBL)	Kimberly Johnson-Miller
Sandia National Laboratories (SNL)	Allen Stanley
Savannah River National Laboratory.....	Saleem Salaymeh
U. S. Nuclear Regulatory Commission (NRC).....	Martha Williams
U. S. Department of Energy (DOE)	James Crabtree
Westinghouse	Rod Likes
Y-12 National Security Complex.....	Sherri Redmon
Member at Large	Obie Amacker, Jr.
Member at Large	Yvonne Ferris
Member at Large	Reuben McGilvary
Member at Large	Bruce Moran
Member at Large	Joseph D. Rivers

Writing Group INMM 5.1 on Measurement Control for the Analytical Laboratory, which developed this standard, had the following members:

- Charles Pietri, Chair
(HITECH Consultants)
- Kenneth Lewis, Secretary
(U.S. Department of Energy, retired)
- Jere T. Bracey
(United States Enrichment Corporation, Paducah Gaseous Diffusion Plant)
- John Clark
(Savannah River Site)
- Yvonne Ferris
(U.S. Department of Energy)
- Steven Goldberg
(U.S. Department of Energy - New Brunswick Laboratory)
- Darryl Jackson
(Los Alamos National Laboratory, retired)
- Peter Mason
(U.S. Department of Energy - New Brunswick Laboratory)
- Melanie May
(U.S. Department of Energy)
- Thomas Pham
(U.S. Nuclear Regulatory Commission)
- Lynne Preston
(U.S. Department of Energy)
- B. Chino Srinivasan
(U.S. Department of Energy - New Brunswick Laboratory)
- Mika Sumi
(Japan Atomic Energy Agency)

American National Standard
for Methods of Nuclear Material Control –

Measurement Control Program – Nuclear Materials Analytical Chemistry Laboratory

0. Introduction

The ability to manage nuclear materials through the principles of material accounting depends on the knowledge of the chemical composition, the concentrations of components, and the associated uncertainty of each measurement of those materials as they are produced, used, shipped, stored, and inventoried. When making the required measurements, the measurement data must be reliable. A comprehensive measurement control program demonstrates the reliability of the measurement data, quantifies the performance of the measurement system, assures that the measurements used in the nuclear industry are suitable for their intended use, and provides for detection and correction of adverse changes.

A laboratory measurement control program should address both the technical and the administrative aspects of measurement processes. The administrative aspects correspond to quality assurance elements, and their implementation establishes quality assurance practices in the measurement control program. Such practices should be traceable to the measurement control requirements of 10 CFR Part 74 and DOE O 470.4, DOE M 470.4-1 and DOE M 470.4-6 (See Annex AA).¹

The goal of any measurement control program is to document and quantify the performance of each analytical measurement system and to provide for detection and correction of adverse changes. The specific needs of each system are determined by identifying the level of performance required and the consequences of using faulty data from that system as part of the material control and accountability program. Additionally, for those facilities that must conduct physical inventories, the measurement control program provides data for establishing the uncertainty (or limit of error) associated with a given inventory difference value. The provisions of this standard provide the basis for evaluating existing practices and modifying those practices, if necessary.

1. Scope and purpose

1.1 Scope. This standard is based on N15.41 (see clause 2, Normative References), which provides the general principles of a measurement control program. This N15.51 standard, although limited to the principal elements of a measurement control program for an analytical chemistry laboratory supporting nuclear fuel cycle activities, has elements that are also applicable to other analytical laboratories.

Measurement control elements specific for bulk measurements (mass and volume) from processes and specific process sampling techniques are not addressed in detail in this standard nor are special analytical techniques such as mass spectrometry and calorimetry. These subjects are treated in other standards of this series.

¹ See annex AA, Bibliography, for the titles and availability information for these publications.