

Technical Information Report

AAMI TIR40: 2009/(R)2016

Sterilization of health care
products—Radiation—
Guidance on dose setting
utilizing a Modified
Method 2

Sterilization of health care products—Radiation— Guidance on dose setting utilizing a Modified Method 2

Approved 3 September 2009 and reaffirmed 18 October 2016 by
Association for the Advancement of Medical Instrumentation

Abstract: This technical information report describes the approach for establishing a sterilization dose utilizing a modification of the fraction positive, incremental dosing method defined in Method 2 of ANSI/AAMI/ISO 11137-2.

Keywords: radiation, dose setting, Modified Method 2

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

This document is not an American National Standard, and the material contained herein is not normative in nature.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every 5 years but at least every 10 years. For a TIR, AAMI consults with a technical committee about 5 years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 301 N Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

Published by

AAMI
4301 N Fairfax Drive, Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2010 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI at 301 N Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 276-0793.

Printed in the United States of America

ISBN 1-57020-366-0

Contents

	Page
Glossary of equivalent standards	iv
Committee representation	vi
Foreword	ix
1 Scope	1
1.1 Inclusions	1
1.2 Exclusions	1
2 Normative References	1
3 Abbreviations, terms, and definitions	2
4 Modified Method 2—Application for method	4
5 Modified Method 2: Dose setting using fraction positive information from incremental dosing to determine an extrapolation factor	4
6 Procedure for Modified Method 2A	5
6.1 General	5
6.2 Stage 1: Select SAL, and obtain samples of product	5
6.3 Stage 2: Perform incremental dose experiment	5
6.3.1 General	5
6.3.2 A and FFP	6
6.3.3 D^*	7
6.3.4 CD^* batch	7
6.4 Stage 3: Perform verification dose experiment	7
6.5 Stage 4: Consideration of results	8
6.6 Stage 5: Establish sterilization dose	8
7 Procedure for Modified Method 2B	9
7.1 General	9
7.2 Stage 1: Select SAL, and obtain samples of product	9
7.3 Stage 2: Perform incremental dose experiment	9
7.3.1 General	9
7.3.2 A and FFP	10
7.3.3 D^*	10
7.3.4 CD^* batch	11
7.4 Stage 3: Perform verification dose experiment	11
7.5 Stage 4: Consideration of results	11
7.6 Stage 5: Establish sterilization dose	12
8 Auditing sterilization dose	12
9 Worked examples	12
9.1 General	12
9.2 Worked example for Modified Method 2A	12
9.2.1 Stage 2: Incremental dose experiment	13
9.2.2 Stage 3: Verification dose experiment	14
9.2.3 Stage 5: Calculation of sterilization dose	15
9.3 Worked example for Modified Method 2B	15
9.3.1 Stage 2: Incremental dose experiment	15
9.3.2 Stage 3: Verification dose experiment	16
9.3.3 Stage 5: Calculation of sterilization dose	17
9.4 Worked example for Modified Method 2A with additional dose required	17
9.4.1 Stage 2: Incremental dose experiment	17
9.4.2 Stage 3: Verification dose experiment	19
9.4.3 Stage 5: Calculation of sterilization dose	20
Bibliography	21

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE: Documents are sorted by international designation.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005 Technical Corrigendum 1 and 2	ANSI/AAMI ES60601-1:2005 ANSI/AAMI ES60601-1:2005/C1:2009 (amdt)	Major technical variations C1 Identical to Corrigendum 1 and 2
IEC 60601-1-2:2007	ANSI/AAMI/IEC 60601-1-2:2007	Identical
IEC 60601-2-2:2009	ANSI/AAMI/IEC 60601-2-2:2009	Identical
IEC 60601-2-4:2002	ANSI/AAMI DF80:2003	Major technical variations
IEC 60601-2-19:2009	ANSI/AAMI/IEC 60601-2-19:2009	Identical
IEC 60601-2-20:2009	ANSI/AAMI/IEC 60601-2-20:2009	Identical
IEC 60601-2-21:2009	ANSI/AAMI/IEC 60601-2-21:2009	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004/(R)2009	Major technical variations
IEC 60601-2-47:2001	ANSI/AAMI EC38:2007	Major technical variations
IEC 60601-2-50:2009	ANSI/AAMI/IEC 60601-2-50:2009	Identical
IEC 80601-2-30:2009	ANSI/AAMI/IEC 80601-2-30:2009	Identical (with inclusion)
IEC 80601-2-58:2008	ANSI/AAMI/IEC 80601-2-58:2008	Identical
IEC/TR 60878:2009	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2009	ANSI/AAMI/IEC TIR62296:2009	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
IEC/TR 80002-1:2009	ANSI/IEC/TR 80002-1:2009	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2004	Identical
ISO 7199:2009	ANSI/AAMI/ISO 7199:2009	Identical
ISO 8637:2004	ANSI/AAMI RD16:2007	Major technical variations
ISO 8638:2004	ANSI/AAMI RD17:2007	Major technical variations
ISO 10993-1:2009	ANSI/AAMI/ISO 10993-1:2009	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:2006	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003/(R)2009	Identical
ISO 10993-4:2002 and Amendment 1:2006	ANSI/AAMI/ISO 10993-4:2002/(R)2009 and Amendment 1:2006/(R)2009	Identical
ISO 10993-5:2009	ANSI/AAMI/ISO 10993-5:2009	Identical
ISO 10993-6:2007	ANSI/AAMI/ISO 10993-6:2007	Identical
ISO 10993-7:2008	ANSI/AAMI/ISO 10993-7:2008	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999/(R)2005	Identical
ISO 10993-10:2002 and Amendment 1:2006	ANSI/AAMI BE78:2002/(R)2008 ANSI/AAMI BE78:2002/A1:2006/(R)2008	Minor technical variations Identical
ISO 10993-11:2006	ANSI/AAMI/ISO 10993-11:2006	Identical
ISO 10993-12:2007	ANSI/AAMI/ISO 10993-12:2007	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999/(R)2004	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001/(R)2006	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000/(R)2006	Identical
ISO 10993-16:1997	ANSI/AAMI/ISO 10993-16:1997/(R)2009	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002/(R)2008	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical
ISO 11135-1:2007	ANSI/AAMI/ISO 11135-1:2007	Identical
ISO/TS 11135-2:2008	ANSI/AAMI/ISO TIR11135-2:2008	Identical

International designation	U.S. designation	Equivalency
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006	Identical
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005	Identical
ISO 11140-3:2007	ANSI/AAMI/ISO 11140-3:2007	Identical
ISO 11140-4:2007	ANSI/AAMI/ISO 11140-4:2007	Identical
ISO 11140-5:2007	ANSI/AAMI/ISO 11140-5:2007	Identical
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006	Identical
ISO 11737-1: 2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:2009	ANSI/AAMI/ISO 11737-2:2009	Identical
ISO 13408-1:2008	ANSI/AAMI/ISO 13408-1:2008	Identical
ISO 13408-2:2003	ANSI/AAMI/ISO 13408-2:2003	Identical
ISO 13408-3:2006	ANSI/AAMI/ISO 13408-3:2006	Identical
ISO 13408-4:2005	ANSI/AAMI/ISO 13408-4:2005	Identical
ISO 13408-5:2006	ANSI/AAMI/ISO 13408-5:2006	Identical
ISO 13408-6:2006	ANSI/AAMI/ISO 13408-6:2006	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003/(R)2009	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003/(R)2008	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003/(R)2008	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998/(R)2008	Identical
ISO 14161:2009	ANSI/AAMI/ISO 14161:2009	Identical
ISO 14708-3:2008	ANSI/AAMI/ISO 14708-3:2008	Identical
ISO 14708-4:2008	ANSI/AAMI/ISO 14708-4:2008	Identical
ISO 14937:2009	ANSI/AAMI/ISO 14937:2009	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2007	ANSI/AAMI/ISO 14971:2007	Identical
ISO 15223-1:2007 and A1:2008	ANSI/AAMI/ISO 15223-1:2007 and A1:2008	Identical
ISO 15225:2000 and A1:2004	ANSI/AAMI/ISO 15225:2000/(R)2006 and A1:2004/(R)2006	Identical
ISO 15674:2009	ANSI/AAMI/ISO 15674:2009	Identical
ISO 15675:2009	ANSI/AAMI/ISO 15675:2009	Identical
ISO 15882:2008	ANSI/AAMI/ISO 15882:2008	Identical
ISO 15883-1:2006	ANSI/AAMI ST15883-1:2009	Major technical variations
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2005	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:2006	ANSI/AAMI/ISO 17665-1:2006	Identical (with inclusions)
ISO/TS 17665-2:2009	ANSI/AAMI/ISO TIR17665-2:2009	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 22442-1:2007	ANSI/AAMI/ISO 22442-1:2007	Identical
ISO 22442-2:2007	ANSI/AAMI/ISO 22442-2:2007	Identical
ISO 22442-3:2007	ANSI/AAMI/ISO 22442-3:2007	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003/(R)2009 and A1:2005/(R)2009	Identical
ISO 25539-2:2008	ANSI/AAMI/ISO 25539-2:2008	Identical
ISO 81060-1:2007	ANSI/AAMI/ISO 81060-1:2007	Identical
ISO 81060-2:2009	ANSI/AAMI/ISO 81060-2:2009	Identical

Committee representation

Association for the Advancement of Medical Instrumentation

Radiation Sterilization Working Group

This technical information report (TIR) was developed by the AAMI Radiation Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Working Group approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Radiation Sterilization Working Group** had the following members:

Cochairs Lisa Foster, Sterigenics International
Stephen C. Yeadon, BS, Kinetic Concepts Inc.

Members Anne F. Booth, MS, Conmed Corporation
Trabue D. Bryans, WuXi AppTec
Harry F. Bushar, PhD (Independent Expert)
Glenn W. Calvert, Becton Dickinson & Company
Gary J. Chilson, Ethox International Inc.
Gary N. Cranston, Consulting & Technical Services/PCS
Elaine Daniell, CR Bard
Douglas D. Davie, Sterilization Validation Services
Darci Diage, Direct Flow Medical
Dave Dion, Cardinal Health (MP&S)
April J. Doering, St. Jude Medical Inc.
Barry P. Fairand, PhD (Independent Expert)
William F. FitzGerald, PE, FitzGerald & Associates Ltd.
Lisa Foster, Sterigenics International
Joyce M. Hansen, JM Hansen & Associates
Thomas L. Hansen, Terumo Medical Corporation
Doug F. Harbrecht, Boston Scientific Corporation
Deborah A. Havlik, Hospira Worldwide Inc.
Katherine E. Hill, 3M Healthcare
Carolyn L. Kinsley, LexaMed
John B. Kowalski, PhD, MicroGAMMA LLC
Byron J. Lambert, PhD, Abbott Laboratories
Michael Larkin, Johnson & Johnson
Jeff Martin, Alcon Laboratories Inc.
John Masefield, Steris Corporation
Joseph M. Mello, Ethide Laboratories Inc.
Russell D. Mills, Zimmer Inc.
Gerry A. O'Dell, MS, Gerry O'Dell Consulting
Kevin O'Hara, MDS Nordion
Dave Parente, NAMSA
Frank Peacock, Jr., Bausch & Lomb Inc.
Rudy M. Pina, Dynatec Scientific Labs Inc.
Manuel Saavedra, Jr., Kimberly-Clark Corporation
Zenius V. Seliokas, Stericon Inc.
Jon Seulean, CaridianBCT Sterilization Services Inc.
William N. Thompson, Covidien
Richard L. Weisman, Fresenius Medical Care Renal Therapies Group
Patrick B. Weixel, FDA/CDRH
James L. Whitby, MA MB FRCP (Independent Expert)
John Andrew Williams, Baxter Healthcare Corporation
Martell Kress Winters, BS SM, Nelson Laboratories Inc.
Stephen C. Yeadon, BS, Kinetic Concepts Inc.

Alternates

John Broad, NAMSA
 David Cardin, Zimmer Inc.
 Charles Cogdill, Boston Scientific Corporation
 John DiCaro, Covidien
 Brian R. Drumheller, CR Bard
 Steven J. Elliott, WuXi AppTec
 Niki Fidopiastis, Sterigenics International
 Thomas Gwise, FDA/CDRH
 Brent Huberty, St. Jude Medical Inc.
 Chris Johnson, Steris Corporation
 Orlando C. Johnson, Hospira Worldwide Inc.
 Jim Kaiser, Bausch & Lomb Inc.
 Bert Kingsbury, Terumo Medical Corporation
 Ezra Koski, A, CaridianBCT Sterilization Services Inc.
 Maggie Ladd, BA BS, Kimberly-Clark Corporation
 Linda Lavelle, Johnson & Johnson
 Reynaldo Lopez, Cardinal Health (MP&S)
 Consuelo McChesney, Alcon Laboratories Inc.
 David Ford McGoldrick, BS, Abbott Laboratories
 Lorra Parliament, Becton Dickinson & Company
 Robert R. Reich, BS MS, LexaMed
 Mark Seybold, Baxter Healthcare Corporation
 Harry L. Shaffer, JM Hansen & Associates
 Larry Talapa, 3M Healthcare
 Jason Voisinet, Ethox International Inc.
 Wendy Wangsgard, PhD, Nelson Laboratories Inc.

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

Cochairs

Victoria M. Hitchins, PhD, FDA/CDRH
 William E. Young, Boston Scientific Corporation

Members

Trabue D. Bryans, WuXi AppTec
 Peter A. Burke, PhD, Steris Corporation
 Nancy Chobin, RN CSPDM, St. Barnabas Healthcare System (Independent Expert)
 Charles Cogdill, Boston Scientific Corporation
 Ramona Conner, RN MSN CNOR, Association of Perioperative Registered Nurses
 Jacqueline Daley, Association for Professionals in Infection Control and Epidemiology
 Kimbrell Darnell, CR Bard
 Lisa Foster, Sterigenics International
 Joel R. Gorski, PhD, NAMSA
 Deborah A. Havlik, Hospira Worldwide Inc.
 Victoria M. Hitchins, PhD, FDA/CDRH
 Danny Hutson, Cardinal Health (MP&S)
 Lois Atkinson Jones, MS (Independent Expert)
 Susan G. Klacik, CCSMC FCS ACE, IAHCMM
 Byron J. Lambert, PhD, Abbott Laboratories
 Colleen Patricia Landers, RN, Canadian Standards Association
 Lisa N. Macdonald, Becton Dickinson & Company
 Jeff Martin, Alcon Laboratories Inc.
 Patrick J. McCormick, PhD, Bausch & Lomb Inc.
 Rainer Newman, Johnson & Johnson
 Janet M. Prust, 3M Healthcare
 Nancy Rakiewicz, Ethox International Inc.
 Michael H. Scholla, Dupont Nonwovens
 Mark Seybold, Baxter Healthcare Corporation
 Andrew Sharavara, PhD, Propper Manufacturing Company Inc.
 Mark N. Smith, Getinge USA
 William N. Thompson, Covidien
 James L. Whitby, MA MB FRCP (Independent Expert)
 Martell Kress Winters, BS SM, Nelson Laboratories Inc.

Alternates

Lloyd Brown, Covidien
Glenn W. Calvert, Becton Dickinson & Company
Dave Dion, Cardinal Health (MP&S)
Steven J. Elliott, WuXi AppTec
Thomas J. Frazar, Johnson & Johnson
Kathy Hoffman, Sterigenics International
Jim Kaiser, Bausch & Lomb Inc.
Joseph J. Lasich, BS, Alcon Laboratories Inc.
Chiu S. Lin, PhD, FDA/CDRH
Natalie Lind, IAHCSSM
Ralph Makinen, Boston Scientific Corporation
Mary S. Mayo, CR Bard
David Ford McGoldrick, BS, Abbott Laboratories
Jerry R. Nelson, PhD, Nelson Laboratories Inc.
Karen Polkinghorne, Dupont Nonwovens
Mike Sadowski, Baxter Healthcare Corporation
John R. Scoville, Jr., Steris Corporation
Jason Voisinet, Ethox International Inc.
Craig A. Wallace, 3M Healthcare
Valerie Welter, Hospira Worldwide Inc.
William E. Young, Boston Scientific Corporation

AAMI also acknowledges the **TIR40 Editing Task Group**, comprising the following members, for its special contribution in the development of this document:

Task group leaders

Deborah A. Havlik, Hospira Worldwide Inc.
Dave Parente, NAMSA

Members

John Broad, NAMSA
Steven J. Elliott, WuXi AppTec
Barry P. Fairand, PhD (Independent Expert)
Niki Fidopiastis, Sterigenics International
John B. Kowalski, PhD, MicroGAMMA LLC
Gerry A. O'Dell, MS, Gerry O'Dell Consulting

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

Foreword

This document is intended to be used in conjunction with ANSI/AAMI/ISO 11137, *Sterilization of health care products—Radiation sterilization* (series). This technical information report (TIR) describes dose-setting methods that may be used to establish the sterilization dose in accordance with one of the two approaches specified in 8.2 (Method 2A) or 8.3 (Method 2B) of ANSI/AAMI/ISO 11137-2:2006.

The basis of the dose-setting methods used in Methods 1 and 2 described in ANSI/AAMI/ISO 11137-2 owes much to the ideas first propounded by Tallentire (Tallentire, 1973; Tallentire, Dwyer, and Ley, 1971; Tallentire and Khan, 1978). Subsequent to these publications, standardized protocols were developed (Davis et al., 1981; Davis, Strawderman, and Whitby, 1984), which formed the basis of the dose-setting methods detailed in the AAMI Recommended Practice for Sterilization by Gamma Radiation (AAMI 1984, 1991).

Methods 2A and 2B of ANSI/AAMI/ISO 11137-2 use data derived from the inactivation of the microbial population in its natural state on product. Using Method 2 can result in a sterilization dose that is significantly less than the sterilization dose established through the use of either Method 1 or Method VD_{max} . Thus, for products that are susceptible to radiation damage, Method 2 could be the most appropriate method of dose establishment. For a Method 2A dose establishment, 640 product units are required, and for a Method 2B dose establishment, 580 are required (see ANSI/AAMI/ISO 11137-2:2006). The method described herein is a modification of Methods 2A and 2B, and uses a reduced number of product items to establish the sterilization dose.

This TIR contains guidelines that are not intended to be absolute or to apply in all circumstances. One should use judgment in applying the information in this TIR.

As used within the context of this document, “should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the technical information report. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation.

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

NOTE—This foreword does not contain provisions of AAMI TIR40:2009, *Sterilization of health care products—Radiation—Guidance on dose setting utilizing a Modified Method 2*, but it does provide important information about the development and intended use of the document.

Sterilization of health care products—Radiation— Guidance on dose setting utilizing a Modified Method 2

1 Scope

1.1 Inclusions

This TIR describes a modification to Methods 2A and 2B of ANSI/AAMI/ISO 11137-2 that reduces the number of incremental doses needed to determine the minimum dose required to achieve a predetermined sterility assurance level (SAL).

This method of sterilization dose establishment may be used to meet the product qualification requirements specified in ANSI/AAMI/ISO 11137-2 when product bioburden is low in number or resistance and has demonstrated historical consistency.

1.2 Exclusions

This method of sterilization dose establishment is not to be used to meet the product qualification requirements specified in ANSI/AAMI/ISO 11137-2 when the product bioburden has not been evaluated.

2 Normative References

The following normative documents contain provisions that, through reference in this text, constitute provisions of this TIR. For dated references, the subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this TIR are encouraged to investigate the possibility of applying the most recent editions of the following normative documents. For undated references, the latest edition of the normative document referred to applies. AAMI maintains a register of currently valid AAMI technical documents.

- a) ANSI/AAMI/ISO 11137-2, *Sterilization of health care products—Radiation—Part 2: Establishing the sterilization dose.*
- b) ANSI/AAMI/ISO 11137-3, *Sterilization of health care products—Radiation—Part 3: Guidance on dosimetric aspects.*

For harmonization and ease of use, portions of the guidance provided in this TIR are written to align with the clauses of ANSI/AAMI/ISO 11137-2.