

ASME MFC-21.2–2010

Measurement of Fluid Flow by Means of Thermal Dispersion Mass Flowmeters

AN AMERICAN NATIONAL STANDARD



ASME MFC-21.2–2010

Measurement of Fluid Flow by Means of Thermal Dispersion Mass Flowmeters

AN AMERICAN NATIONAL STANDARD



Three Park Avenue • New York, NY • 10016 USA

Date of Issuance: January 10, 2011

This Standard will be revised when the Society approves the issuance of a new edition. There will be no addenda issued to this edition.

ASME issues written replies to inquiries concerning interpretations of technical aspects of this document. Periodically certain actions of the ASME MFC Committee may be published as Cases. Cases and interpretations are published on the ASME Web site under the Committee Pages at <http://cstools.asme.org> as they are issued.

ASME is the registered trademark of The American Society of Mechanical Engineers.

This code or standard was developed under procedures accredited as meeting the criteria for American National Standards. The Standards Committee that approved the code or standard was balanced to assure that individuals from competent and concerned interests have had an opportunity to participate. The proposed code or standard was made available for public review and comment that provides an opportunity for additional public input from industry, academia, regulatory agencies, and the public-at-large.

ASME does not “approve,” “rate,” or “endorse” any item, construction, proprietary device, or activity.

ASME does not take any position with respect to the validity of any patent rights asserted in connection with any items mentioned in this document, and does not undertake to insure anyone utilizing a standard against liability for infringement of any applicable letters patent, nor assume any such liability. Users of a code or standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, is entirely their own responsibility.

Participation by federal agency representative(s) or person(s) affiliated with industry is not to be interpreted as government or industry endorsement of this code or standard.

ASME accepts responsibility for only those interpretations of this document issued in accordance with the established ASME procedures and policies, which precludes the issuance of interpretations by individuals.

No part of this document may be reproduced in any form,
in an electronic retrieval system or otherwise,
without the prior written permission of the publisher.

The American Society of Mechanical Engineers
Three Park Avenue, New York, NY 10016-5990

Copyright © 2011 by
THE AMERICAN SOCIETY OF MECHANICAL ENGINEERS
All rights reserved
Printed in U.S.A.

CONTENTS

Foreword		iv
Committee Roster		v
Correspondence With the MFC Committee		vi
1	Scope	1
2	Terminology and Symbols	1
3	General Description	5
4	Principle of Operation	8
5	Guidelines for Flowmeter Selection	12
6	Guidelines for Installation and Applications	14
7	Inspection and Compliance	18
8	Safety	18
9	References	19
Figures		
3.1-1	The Major Components of Two Configurations of Thermal Dispersion Mass Flowmeters	6
3.2-1	Flow Sensor of Thermal Dispersion Mass Flowmeters	7
3.3-1	Two Modes of Flow Sensor Operation	9
4-1	Principle of Operation	9
Tables		
2.3-1	Symbols	2
2.4-1	Abbreviations	4
6.1.4-1	Straight Pipe Length Requirements for an In-Line Flowmeter With a Built-In Flow Conditioner	16
6.1.4-2	Straight Pipe Length Requirements for Insertion Flowmeters With No Flow Conditioning	16
Mandatory Appendix		
I	Flow Calibration	21
Nonmandatory Appendices		
A	In-Line Flowmeter Sizing and Permanent Pressure Loss	25
B	Accuracy and Uncertainty	27

FOREWORD

Thermal dispersion mass flowmeters comprise a family of instruments for the measurement of the total mass flow rate of a fluid, primarily gases, flowing through closed conduits.

The operation of thermal dispersion mass flowmeters is attributed to L.V. King who, in 1914 [1], published his famous King's Law revealing how a heated wire immersed in a fluid flow measures the mass velocity at a point in the flow. King called his instrument a "hot-wire anemometer." However, it was not until the 1960s and 1970s that industrial-grade thermal dispersion mass flowmeters finally emerged.

This Standard covers the thermal dispersion type of thermal mass flowmeter. A companion standard, ASME MFC 21.1, Measurement of Fluid Flow by Means of Capillary Tube Thermal Mass Flowmeters and Controllers, covers the other most commonly used type of thermal mass flowmeter. Both types measure fluid mass flow rate by means of the heat convected from a heated surface to the flowing fluid. In the case of the thermal dispersion, or immersible, type of flowmeter, the heat is transferred to the boundary layer of the fluid flowing over the heated surface. In the case of the capillary tube type of flowmeter the heat is transferred to the bulk of the fluid flowing through a small heated capillary tube. The principles of operation of the two types are both thermal in nature, but are so substantially different that two separate standards are required. Additionally, their applications are much different. Thermal dispersion flowmeters are commonly used for general industrial gas-flow applications in pipes and ducts, whereas capillary tube flowmeters are primarily used for smaller flows of clean gases in tubes.

Suggestions for improvement of this Standard will be welcomed. They should be sent to The American Society of Mechanical Engineers; Attn: Secretary, MFC Standards Committee; Three Park Avenue; New York, NY 10016-5990.

This Standard was approved as an American National Standard on June 24, 2010.

ASME MFC COMMITTEE

Measurement of Fluid Flow in Closed Conduits

(The following is the roster of the Committee at the time of approval of this Standard.)

STANDARDS COMMITTEE OFFICERS

R. J. DeBoom, *Chair*
Z. D. Husain, *Vice Chair*
D. C. Wyatt, *Vice Chair*
C. J. Gomez, *Secretary*

STANDARDS COMMITTEE PERSONNEL

C. J. Blechinger, *Honorary Member*, Consultant
R. M. Bough, Rolls-Royce Motor Cars
M. S. Carter, Flow Systems, Inc.
G. P. Corpron, *Honorary Member*, Consultant
R. J. DeBoom, Consultant
D. Faber, *Contributing Member*, Badger Meter, Inc.
C. J. Gomez, The American Society of Mechanical Engineers
F. D. Goodson, Emerson Process
Z. D. Husain, Chevron Corp.
C. G. Langford, *Honorary Member*, Consultant
T. O. Maginnis, Consultant
W. M. Mattar, Invensys/Foxboro Co.

G. E. Mattingly, The Catholic University of America
D. R. Mesnard, Consultant
R. W. Miller, *Honorary Member*, R. W. Miller & Associates, Inc.
A. Quraishi, American Gas Association
W. Seidl, Colorado Engineering Experiment Station, Inc.
D. W. Spitzer, *Contributing Member*, Spitzer and Boyes LLC
R. N. Steven, Colorado Engineering Experiment Station, Inc.
T. M. Kegel, *Alternate*, Colorado Engineering Experiment Station, Inc.
J. H. Vignos, *Honorary Member*, Consultant
D. E. Wiklund, Rosemount, Inc.
J. D. Wright, *Contributing Member*, NIST
D. C. Wyatt, Wyatt Engineering

SUBCOMMITTEE 21 — THERMAL MASS FLOWMETERS

R. J. DeBoom, *Chair*, Consultant
T. O. Maginnis, *Vice Chair*, Consultant
Z. D. Husain, Chevron Corp.

J. G. Olin, Sierra Instruments, Inc.
B. K. Rao, Consultant

CORRESPONDENCE WITH THE MFC COMMITTEE

General. ASME Standards are developed and maintained with the intent to represent the consensus of concerned interests. As such, users of this Standard may interact with the Committee by requesting interpretations, proposing revisions, and attending Committee meetings. Correspondence should be addressed to:

Secretary, MFC Standards Committee
The American Society of Mechanical Engineers
Three Park Avenue
New York, NY 10016-5990
<http://go.asme.org/Inquiry>

Proposing Revisions. Revisions are made periodically to the Standard to incorporate changes that appear necessary or desirable, as demonstrated by the experience gained from the application of the Standard. Approved revisions will be published periodically.

The Committee welcomes proposals for revisions to this Standard. Such proposals should be as specific as possible, citing the paragraph number(s), the proposed wording, and a detailed description of the reasons for the proposal, including any pertinent documentation.

Proposing a Case. Cases may be issued for the purpose of providing alternative rules when justified, to permit early implementation of an approved revision when the need is urgent, or to provide rules not covered by existing provisions. Cases are effective immediately upon ASME approval and shall be posted on the ASME Committee Web page.

Requests for Cases shall provide a Statement of Need and Background Information. The request should identify the Standard, the paragraph, figure or table number(s), and be written as a Question and Reply in the same format as existing Cases. Requests for Cases should also indicate the applicable edition(s) of the Standard to which the proposed Case applies.

Interpretations. Upon request, the MFC Committee will render an interpretation of any requirement of the Standard. Interpretations can only be rendered in response to a written request sent to the Secretary of the MFC Standards Committee.

The request for interpretation should be clear and unambiguous. It is further recommended that the inquirer submit his/her request in the following format:

Subject: Cite the applicable paragraph number(s) and the topic of the inquiry.
Edition: Cite the applicable edition of the Standard for which the interpretation is being requested.
Question: Phrase the question as a request for an interpretation of a specific requirement suitable for general understanding and use, not as a request for an approval of a proprietary design or situation. The inquirer may also include any plans or drawings, that are necessary to explain the question; however, they should not contain proprietary names or information.

Requests that are not in this format may be rewritten in the appropriate format by the Committee prior to being answered, which may inadvertently change the intent of the original request.

ASME procedures provide for reconsideration of any interpretation when or if additional information that might affect an interpretation is available. Further, persons aggrieved by an interpretation may appeal to the cognizant ASME Committee or Subcommittee. ASME does not "approve," "certify," "rate," or "endorse" any item, construction, proprietary device, or activity.

Attending Committee Meetings. The MFC Standards Committee regularly holds meetings that are open to the public. Persons wishing to attend any meeting should contact the Secretary of the MFC Standards Committee.

MEASUREMENT OF FLUID FLOW BY MEANS OF THERMAL DISPERSION MASS FLOWMETERS

1 SCOPE

This Standard establishes common terminology and gives guidelines for the quality, description, principle of operation, selection, installation, and flow calibration of thermal dispersion flowmeters for the measurement of the mass flow rate, and to a lesser extent, the volumetric flow rate, of the flow of a fluid in a closed conduit. Multivariable versions additionally measure fluid temperature. Thermal dispersion mass flowmeters are applicable to the flow of single-phase pure gases and gas mixtures of known composition and, less commonly, to single-phase liquids of known composition. Companion standard ASME MFC-21.1 covers capillary tube type thermal mass flowmeters and controllers.

2 TERMINOLOGY AND SYMBOLS

(a) Paragraph 2.1 lists definitions from ASME MFC-1M used in ASME MFC-21.2.

(b) Paragraph 2.2 lists definitions specific to this Standard.

(c) Paragraph 2.3 lists symbols (see Table 2.3-1) used in this Standard (see notes and superscripts).

(d) Paragraph 2.4 lists abbreviations (see Table 2.4-1) used in this Standard.

2.1 Definitions Copied From ASME MFC-1M

accuracy: the degree of freedom from error; the degree of conformity of the indicated value to the true value of the measured quantity.

calibration: the experimental determination of the relationship between the quantity being measured and the device that measures it, usually by comparison with a standard. Also, the act of adjusting the output of a device to bring it to a desired value, within a specified tolerance, for a particular value of the input.

cavitation: the implosion of vapor bubbles formed after flashing when the local pressure rises above the vapor pressure of a liquid. (See also *flashing*.)

flashing: the formation of vapor bubbles in a liquid when the local pressure falls to or below the vapor pressure of the liquid, often due to local lowering of pressure because of an increase in the liquid velocity. (See also *cavitation*.)

flow profile: graphic representation of the velocity distribution.

fully developed velocity distribution: a velocity distribution, in a straight length of pipe, that has zero radial and azimuthal fluid velocity components and an axisymmetric axial velocity profile that is independent of axial position along the pipe.

rangeability (turndown): flowmeter rangeability is the ratio of the maximum to minimum flow rates or Reynolds number in the range over which the meter meets a specified uncertainty (accuracy).

repeatability (qualitative): the closeness of agreement among a series of results obtained with the same method on identical test material, under the same conditions (same operator, same apparatus, same laboratory, and short intervals of time). [See also *repeatability (quantitative)*.]

repeatability (quantitative): the value below which the absolute difference between any two single test results obtained under same conditions may be expected to lie with a specified probability. In the absence of other indications, the probability is 95%. [See also *repeatability (qualitative)*.]

reproducibility (quantitative): the closeness of agreement between results obtained when the conditions of measurement differ; for example, with respect to different test apparatus, operators, facilities, time intervals, etc.

swirling flow: flow that has axial and circumferential velocity components.

transmitter (secondary device): electronic system providing the drive and transforming the signals from the flow sensor to give output(s) of measured and inferred parameters; it also provides corrections derived from parameters such as temperature.

uncertainty interval, u : an estimate of the error band, centered about the measurement, within which the true value must fall with a specified probability.

2.2 Definitions Specific to This Document

base conditions: the conditions of temperature and pressure to which measured volumes are to be corrected (same as *Reference or Standard Conditions*).